Certificate of Sterility

Product:	Gamma Wipe 67
Item#:	GW67.ST.25
Lot #:	0810819
Irradiation Run #:	1445123
Date of Manufacture:	January 22, 2015
Expiration Date:	01/2018
Manufacturing Site:	Whitsett, North Carolina

This certifies that the above product has been sterilized with Cobalt 60 gamma radiation and has been dosimetrically released as sterile. The specified minimum dose has been delivered to the product. This is in accordance with procedures outlined in ANSI/AAMI/ISO 11137-2:2013: Sterilization of health care products – Radiation – Part 2: Establishing the sterilization dose, to a Sterility Assurance Level of 10-6.

Minimum Specified Dose: 25.0 kGy Minimum Delivered Dose: 28.5 kGy Maximum Specified Dose: 42.0 kGy Maximum Delivered Dose: 36.0 kGy

Reviewed and approved by: Dewayne Steele

DeWayne Steele

Quality Assurance Specialist

Approval Date: January 26, 2015





USA: London: Singapore: Tokvo: Berkshire

800.242.7000/413.528.2602 44. (0).870.757.2877 65.6252.4313 81.3.5827.2380