
Certificate of Sterility

Product:	Gamma Wipe SuperSorb
Item#:	GWSS.ST.16
Lot #:	0808007
Irradiation Run #:	1233946
Date of Manufacture:	January 16, 2014
Expiration Date:	01/2017
Manufacturing Site:	Whitsett, North Carolina

This certifies that the above product has been sterilized with Cobalt 60 gamma radiation and has been dosimetrically released as sterile. The specified minimum dose has been delivered to the product. This is in accordance with procedures outlined in ANSI/AAMI/ISO 11137-2:2013: Sterilization of health care products – Radiation – Part 2: Establishing the sterilization dose, to a Sterility Assurance Level of 10^{-6} .

Minimum Specified Dose: <u>25.0 kGy</u>	Minimum Delivered Dose: <u>30.6 kGy</u>
Maximum Specified Dose: <u>50.0 kGy</u>	Maximum Delivered Dose: <u>40.5 kGy</u>

Reviewed and approved by:



Steve Shoffner
Quality Assurance Administrator
Approval Date: January 29, 2014

Berkshire



Berkshire
USA: 800.242.7000/413.528.2602
London: 44. (0).870.757.2877
Singapore: 65.6252.4313
Tokyo: 81.3.5827.2380

www.berkshire.com
