

---

---

## Certificate of Sterility

<b>Product:</b>	Gamma Wipe SuperSorb
<b>Catalog No.:</b>	GWSS.ST.16
<b>Lot #:</b>	0810757
<b>Irradiation Run #:</b>	1441185
<b>Date of Manufacture:</b>	January 13, 2015
<b>Expiration Date:</b>	01/2018
<b>Manufacturing Site:</b>	Whitsett, North Carolina

This certifies that the above product has been sterilized with Cobalt 60 gamma radiation and has been dosimetrically released as sterile. The specified minimum dose has been delivered to the product. This is in accordance with procedures outlined in ANSI/AAMI/ISO 11137:2006: Sterilization of health care products – Radiation – Part 2: Establishing the sterilization dose, to a Sterility Assurance Level of  $10^{-6}$  and AAMI TIR33:2005: Sterilization of health care products - Radiation – Substantiation of a selected sterilization dose – Method VD<sub>max</sub>.

Minimum Specified Dose: <u>25.0 kGy</u>	Minimum Delivered Dose: <u>27.8 kGy</u>
Maximum Specified Dose: <u>50.0 kGy</u>	Maximum Delivered Dose: <u>41.5 kGy</u>

Reviewed and approved by:

*DeWayne Steele*

DeWayne Steele

Quality Assurance Specialist

Approval Date: January 22, 2015

# Berkshire



**Berkshire**  
USA: 800-242-7000/413-528-2602  
London: 44-(0) 870-757-2877  
Singapore: 65-6252-4313  
Tokyo: 81-3-5827-2380

[www.berkshire.com](http://www.berkshire.com)

---

---