
Certificate of Sterility

Product:	Gamma Wipe CapSure VP
Catalog No.:	GWCPSPV.ST.25
Lot #:	0800155
Irradiation Run #:	638966
Date of Manufacture:	April 26, 2011
Expiration Date:	April 2014
Manufacturing Site:	Whitsett, North Carolina

This certifies that the above product has been sterilized with Cobalt 60 gamma radiation and has been dosimetrically released as sterile. The specified minimum dose has been delivered to the product. This is in accordance with procedures outlined in ANSI/AAMI/ISO 11137:2006: Sterilization of health care products – Radiation – Part 2: Establishing the sterilization dose, to a Sterility Assurance Level of 10^{-6} and AAMI TIR33:2005: Sterilization of health care products - Radiation – Substantiation of a selected sterilization dose – Method VD_{max}.

Minimum Specified Dose: 25.0 kGy Minimum Delivered Dose: 29.4 kGy
Maximum Specified Dose: 50.0 kGy Maximum Delivered Dose: 45.2 kGy

Reviewed and approved by:

Lynn Stanard

Lynn Stanard
Quality Manager

Approval Date: April 27, 2011

Berkshire



Berkshire

USA: 800-242-7000/413-528-2602
London: 44-870-757-2877
Singapore: 65-252-4313
Tokyo: 81-3-5827-2380