
Certificate of Sterility

Product:	Gamma Wipe 300
Catalog No.:	GW300.ST.25
Lot #:	0803309
Irradiation Run #:	878888
Date of Manufacture:	May 21, 2011
Expiration Date:	May 2015
Manufacturing Site:	Whitsett, North Carolina

This certifies that the above product has been sterilized with Cobalt 60 gamma radiation and has been dosimetrically released as sterile. The specified minimum dose has been delivered to the product. This is in accordance with procedures outlined in ANSI/AAMI/ISO 11137:2006: Sterilization of health care products – Radiation – Part 2: Establishing the sterilization dose, to a Sterility Assurance Level of 10^{-6} and AAMI TIR33:2005: Sterilization of health care products - Radiation – Substantiation of a selected sterilization dose – Method VD_{max}.

Minimum Specified Dose: 25.0 kGy
Maximum Specified Dose: 50.0 kGy

Minimum Delivered Dose: 27.9 kGy
Maximum Delivered Dose: 38.7 kGy

Reviewed and approved by:



Steve Shoffner

Quality Assurance Admin

Approval Date: May 31, 2012

Berkshire



Berkshire

USA: 800-242-7000/413-528-2602
London: 44-870-757-2877
Singapore: 65-252-4313
Tokyo: 81-3-5827-2380